

<b>CODE</b>	<b>DESCRIPTION: TOP-KIT TYVEK OVERALL + HOOD + BOOTS – STER. BETA</b>
S1770694	<ol style="list-style-type: none"> <li>1. Insert components according to the following order: BOOT – OVERALL – HOOD.</li> <li>2. Labelling and DOUBLE BAG.</li> <li>3. 20 pieces per box</li> <li>4. Sterilization BETA radiation.</li> </ol>

**Pict. model and dimensions**

**COD. 11015566**

	S	M	L	XL	XXL
A	110	120	130	140	150
B	84	86	89	92	95
C	162	167	171	178	183
D	16	16	16	16	18
E	26	26	27	27	27

**COD. 12105004**

**COD. 11015010**

+/- 2 cm  
+/- 1 cm  
+/- 3 cm  
+/- 1 cm  
+/- 1 cm

Bag label

Box label

**STERILINE DISPOSABLE**

**SET STERILE IN TYVEK**  
COMPOSTO DA:

- 1 TOP/TUTA SALDATA TG. XL
- 1 TOP/CALZARE SALDATO SUOLA ANTISCIVOLO MISURA GRANDE
- 1 TOP/CAPPUCCIO CON BOTTONI

**STERILE BETA**

RED / ORANGE COLOR = IRRADIATED  
UNI EN ISO 11140-1

**COD. S1770694**

Lotto  
Ch.-B.  
Batch no.  
No. de lot

Data ster.  
Steril.-Dat.  
Steril. date

Scadenza  
Expiry  
Peremption  
Verfall

**INDUTEX S.p.A.**  
Via San Francesco n°8/10 - 20011 CORBETTA - MI  
Tel. 02/972.387 11 Fax 02/972.387 99

**STERILINE DISPOSABLE**

18634058678314

**SET STERILE IN TYVEK**  
COMPOSTO DA:

- 1 TOP/TUTA SALDATA TG. XL
- 1 TOP/CALZARE SALDATO SUOLA ANTISCIVOLO MISURA GRANDE
- 1 TOP/CAPPUCCIO CON BOTTONI

**STERILE BETA**

RED / ORANGE COLOR = IRRADIATED  
UNI EN ISO 11140-1

**COD. S1770694**    **25 PZ.**

Lotto  
Ch.-B.  
Batch no.  
No. de lot

Data ster.  
Steril.-Dat.  
Steril. date

Scadenza  
Expiry  
Peremption  
Verfall

**INDUTEX S.p.A.**  
Via San Francesco n°8/10 - 20011 CORBETTA - MI  
Tel. 02/972.387 11 Fax 02/972.387 99

Emissione: RS

Verifica: PRO

Verifica: CQ

Approv: AU  
Mod-09.06 Vers.1 Del 26.01.2000