


PDS No. 63219x	PRODUCT DATA SHEET				Page 1 of 1	
Revision 02	Petri Dish, without Vents, 94 x 15 mm				 greiner bio-one	
	Greiner Item-No. 63219x					
Valid for Item-No.:	632190	632191 (sterile)				

1.	Description / Specification	
1.1	Description	Petri Dish without vents, 94 x 15 mm 632190: standard design, without logo 632191: standard design, without logo, sterile
1.2	Dimensions	See Customer Drawing Total weight: 13,0-14,5 g
1.3	Volume	Max. volume: 80 ml
1.4	Material / Resin	Dish and lid: PS (Polystyrene), free of heavy metal
1.5	Colour	Dish and lid: clear
1.6	Sterilisation	632190: no 632191: SAL 10 <sup>-3</sup>
1.7	Quality Control	- <u>Raw Material-Control</u> : physical testing - <u>Product-Control</u> : testing of attributive and variable characteristics in accordance with the valid specification
1.8	Other Information	For single use only

2.	Features	
2.1	Basic features	Hydrophobic
2.2	Temperature range	-20°C to +60°C
2.3	Autoclavability	No
2.4	Centrifugation, max. RCF	N/A
2.5	Chemical Resistance	See homepage: <a href="https://www.gbo.com/en_INT/know-how-services/download-center.html">https://www.gbo.com/en_INT/know-how-services/download-center.html</a>
2.6	Shelf life	632190: n/a 632191: 5 years after month of production
2.7	Other Information	-

3.	Packaging	
3.1	Pieces / Bag	20
3.2	Pieces / Box	480
3.3	Lot-No.	F YY MM XXX (manufacturing facility, year, month, consecutive SAP-No.)
3.4	Other Information	-

4.	Other Information	
	-	

Data Sheet subject to change without notice!

Prior Issue	Drawn	Approved	Released	<b>CONFIDENTIAL:</b> Information contained in this document or drawing is confidential and proprietary to Greiner Bio-One GmbH. This document may not be reproduced for any reason without written permission from Greiner Bio-One GmbH. All rights of design, invention, and copyright are reserved.
Revision 01	Date 2 November 2015	Date 3 November 2015	Date 3 November 2015	
Date 13.05.2015	Name S. Kaelberer	Name A. Lang	Name Dr. A. Ganser	

DISCLAIMER: The description of a certain product can only be considered as a guidance, because its performance ultimately depends on what the product is used for. Very often performance studies are indispensable.